

| | | |
|--|--|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 13D0869535 | (X3) Date Survey Completed 06/22/2022 |
| Name of Provider or Supplier Southeast Idaho Family Practice | Street Address, City, State 2775 Channing Way, Idaho Falls, ID | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|---|
| D5477 | <p>CONTROL PROCEDURES CFR(s): 493.1256(e)(4)(g)</p> <p>(e) For reagent, media, and supply checks, the laboratory must do the following: (e) (4) Before, or concurrent with the initial use-- (e)(4)(i) Check each batch of media for sterility if sterility is required for testing; (e)(4)(ii) Check each batch of media for its ability to support growth and, as appropriate, select or inhibit specific organisms or produce a biochemical response; and (e)(4)(iii) Document the physical characteristics of the media when compromised and report any deterioration in the media to the manufacturer. (g) The laboratory must document all control procedures performed.</p> <p>This STANDARD is not met as evidenced by: Based on a direct observation of microbiology media, review of control logs and an interview with the technical consultant (TC) on 6/22/2022, the laboratory failed to perform quality control (QC) on group A beta strep agar. The findings include: 1. A direct observation during the laboratory tour identified Hardy group A beta strep agar lot number 504712 expiration 2022-07-11. 2. A review of the laboratory's Bacitracin and group A beta strep agar QC log identified that the laboratory failed to perform QC on group A beta strep agar lot number 504712 to show the ability of the media to support growth and the ability to inhibit growth of specific organisms. 3. An interview with the TC on 6/22/2022 at 3:22 pm confirmed that the laboratory failed to perform QC on group A beta strep agar to show the ability to support growth and inhibit specific organisms for lot number 504712. 4. The laboratory reports performing 165 group A beta streptococcus cultures annually.</p> |