Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 13D0917694	**(X3) Date Survey Completed** 10/28/2025
Name of Provider or Supplier High Valley Dermatology And	**Street Address, City, State** 703 Rigby Lake Dr, Rigby, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D5209	PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235 As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency. This STANDARD is not met as evidenced by: Based on a review of the Centers for Medicare and Medicaid Services (CMS) 209 personnel form, a lack of documentation and an interview with the laboratory lead on 10/28/2025, the laboratory failed to follow written procedures to assess testing personnel's training and competency for potassium hydroxide (KOH) examinations in 2024 and 2025. The findings include: 1. A lack of training records identified one testing personnel (TP) listed on the CMS 209 hired on 1/18/2024 that failed to have documentation of initial training for KOH examinations. 2. A lack of competency assessment records identified one of two TP listed on the CMS 209 hired since the last inspection on 10/27/2023 that failed to have documentation of a six month competency assessment for KOH examinations. 3. A lack of competency assessment records identified one of two TP listed on the CMS 209 that failed to have an annual competency assessment for KOH examinations in 2024 and two of two TP listed on the CMS 209 that failed to have competency assessments for KOH examinations in 2025. 4. An interview with the laboratory lead on 10/28/2025 at 10:32 am confirmed the above findings. 5. The laboratory reports performing 40 KOH examinations annually.