

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D0926381	(X3) Date Survey Completed 08/26/2025
Name of Provider or Supplier Community Care Of Channing Way	Street Address, City, State 2725 Channing Way, Idaho Falls, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5413	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b)</p> <p>(b) The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (b)(1) Water quality. (b)(2) Temperature. (b)(3) Humidity. (b)(4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.</p> <p>This STANDARD is not met as evidenced by: Based on a lack of laboratory humidity records and an interview with the laboratory director on 8/26/2025, the laboratory failed to establish and monitor the humidity for the Sysmex XN-L 430 complete blood count (CBC) testing since its installation for patient testing in November 2024. The findings include: 1. A lack of the laboratory humidity records identified the laboratory failed to establish and monitor humidity for CBC testing per the Sysmex XN-L 430 manufacturer requirements. 2. An interview with the laboratory director on 8/26/2025 at 2:30 pm confirmed that the laboratory failed to monitor humidity since November 2024. 3. The laboratory reports performing 7242 hematology tests on the Sysmex XN-L 430 annually.</p>