

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D0933122	(X3) Date Survey Completed 12/04/2023
Name of Provider or Supplier Scholes Dermatology	Street Address, City, State 526 Shoup Ave W Ste A, Twin Falls, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5473	<p>CONTROL PROCEDURES CFR(s): 493.1256(e)(2)(g)</p> <p>(e) For reagent, media, and supply checks, the laboratory must do the following: (e) (2) Each day of use (unless otherwise specified in this subpart), test staining materials for intended reactivity to ensure predictable staining characteristics. Control materials for both positive and negative reactivity must be included, as appropriate. (g) The laboratory must document all control procedures performed.</p> <p>This STANDARD is not met as evidenced by: Based on a review of patient reports, the laboratory Hematoxylin and Eosin (H&E) stain quality control (QC) calendar and an interview with the laboratory director (LD) on 12/4/2023, the laboratory failed to document QC each day of patient testing. The findings include: 1. A random review of patient reports, slides and the corresponding H&E QC identified that the laboratory failed document H&E stain quality on 6/21/23 and 1/24/2023. 2. An interview with the LD on 12/4/2023 at 9:41 am confirmed the above finding. 3. The laboratory reports performing 1200 dermatopathology tests annually.</p>