

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D0935092	(X3) Date Survey Completed 02/27/2018
Name of Provider or Supplier Gem State Dermatology	Street Address, City, State 388 E Parkcenter Blvd, Boise, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review and an interview with the laboratory medical assistant, the laboratory failed to verify the accuracy of potassium hydroxide (KOH) and Tzank smears at least twice annually since the last survey on April 21, 2016. Findings: 1. A record review revealed the laboratory failed to document the accuracy of KOH and Tzank smears, at least semiannually, for the detection of fungal elements and herpes cells since the last survey on April 21, 2016. 2. An interview on February 27, 2018 at 10:00 AM, with the laboratory medical assistant, confirmed the laboratory failed to document the accuracy of KOH and Tzank smears.</p>