

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D0935096	(X3) Date Survey Completed 06/22/2022
Name of Provider or Supplier Primary Care Specialists	Street Address, City, State 110 Vista Dr, Pocatello, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5805	<p>TEST REPORT CFR(s): 493.1291(c)</p> <p>The test report must indicate the following: (c)(1) For positive patient identification, either the patient's name and identification number, or a unique patient identifier and identification number. (c)(2) The name and address of the laboratory location where the test was performed. (c)(3) The test report date. (c)(4) The test performed. (c)(5) Specimen source, when appropriate. (c)(6) The test result and, if applicable, the units of measurement or interpretation, or both. (c)(7) Any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.</p> <p>This STANDARD is not met as evidenced by: Based on a review of patient laboratory reports and an interview with the laboratory lead on 6/22/2022, the laboratory failed to indicate the name and address of the performing laboratory for wet mount preparations and urine sediment examinations. The findings include: 1. A review of patient laboratory test reports for wet mount preparations and urine sediment examinations identified that the laboratory failed to indicate the name and address of the performing laboratory since implementing their new electronic medial record system in April 2021. 2. An interview with the laboratory lead on 6/22/2022 at 10:07 am confirmed that the name and address of the performing laboratory for wet mount preparations and urine sediment examinations are not indicated on the test report. 3. The laboratory reports performing 956 wet mount preparations and urine sediment examinations annually.</p>
D5807	<p>TEST REPORT CFR(s): 493.1291(d)</p> <p>Pertinent "reference intervals" or "normal" values, as determined by the laboratory performing the tests, must be available to the authorized person who ordered the tests</p>

and, if applicable, the individual responsible for using the test results.

This STANDARD is not met as evidenced by:

Based on a review of patient laboratory reports and an interview with the laboratory lead on 6/22/2022, the laboratory failed to indicate the normal values for wet mount preparations and urine sediment examinations. The findings include: 1. A review of patient laboratory test reports for wet mount preparations and urine sediment examinations identified that the laboratory failed to indicate the normal values since implementing their new electronic medial record system in April 2021. 2. An interview with the laboratory lead on 6/22/2022 at 10:07 am confirmed that the normal values for wet mount preparations and urine sediment examinations are not indicated on the test report. 3. The laboratory reports performing 956 wet mount preparations and urine sediment examinations annually.