

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D0935096	(X3) Date Survey Completed 05/14/2024
Name of Provider or Supplier Primary Care Specialists	Street Address, City, State 110 Vista Dr, Pocatello, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2009	<p>TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(1)</p> <p>The individual testing or examining the samples and the laboratory director must attest to the routine integration of the samples into the patient workload using the laboratory's routine methods.</p> <p>This STANDARD is not met as evidenced by: Based on a review of proficiency testing (PT) records from the American Proficiency Institute (API) and an interview with testing personnel 1 (TP1) on 5/14/2024, the laboratory failed to sign PT attestations in 2022, 2023 and 2024. The findings include: 1. A review of PT records from API for event three in 2022 identified that the testing personnel failed to sign the attestation statement for potassium hydroxide (KOH) slide examination, wet mount slide examination and urine sediment testing. 2. A review of PT records from API for events one and three in 2023 identified that the laboratory director and testing personnel failed to sign the attestation statements for KOH slide examination, wet mount slide examination and urine sediment testing. 3. A review of PT records from API for event two in 2023 identified that the testing personnel failed to sign the attestation statement for KOH slide examination, wet mount slide examination and urine sediment testing. 4. A review of PT records from API for event one in 2024 identified that the laboratory director and testing personnel failed to sign the attestation statements for KOH slide examination, wet mount slide examination and urine sediment testing 5. An interview with TP1 on 5/14/2024 at 1:23 pm confirmed that the laboratory director and testing personnel failed to sign attestation statements for the above events. 6. The laboratory reports performing 144 KOH slide examination, wet mount slide examination and urine sediment tests annually.</p>