

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  13D0966533	<b>(X3) Date Survey Completed</b>  12/12/2025
<b>Name of Provider or Supplier</b>  North Idaho Dermatology Pa	<b>Street Address, City, State</b>  2199 N Merritt Creek Loop, Coeur D'Alene, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on a direct observation and an interview with the laboratory manager on 12/12/2025, the laboratory failed to discontinue the use of expired potassium hydroxide (KOH) reagent prior to its expiration. The findings include: 1. A direct observation of the laboratory's reagents on 12/12/2025 identified that the laboratory failed to discontinue the use of Delasco 10% KOH w/DMSO, lot K2298A, expiration 9/30/2025 prior to the expiration. 2. An interview with the laboratory manager on 12/12/2025 at 10:30 am confirmed the use of the expired KOH reagent. 3. The laboratory reports performing 140 KOH slide examinations yearly.</p>