

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D0980387	(X3) Date Survey Completed 03/06/2025
Name of Provider or Supplier Ip Boise 1	Street Address, City, State 8660 W Emerald St #102, Boise, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5413	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b)</p> <p>(b) The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (b)(1) Water quality. (b)(2) Temperature. (b)(3) Humidity. (b)(4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.</p> <p>This STANDARD is not met as evidenced by: Based on a direct observation, review of manufacturer's instructions, and interview with the technical supervisor on 3/6/2025, the laboratory failed to monitor the room temperature in the supply storage room where testing reagents were stored. The findings include: 1. A direct observation on 3/6/2025 at 1:50 pm identified the following reagents in the storage room for which the laboratory failed to monitor the room temperature: a. Beckman Coulter DxH 500 cleaner; storage temperature 4-25 C b. Beckman Coulter DxH 500 lyse; storage temperature 4-25 C c. Polymedco Sed-Chek 2; storage temperature 18-35 C d. Beckman Coulter AU ISE selectivity check; storage temperature 2-25 C e. Beckman Coulter AU REF solution; storage temperature 15-25 C f. Stago STA cleaner solution; storage temperature 2-25 C 2. An interview with the technical supervisor on 3/6/2025 at 1:52 pm confirmed that they didn't monitor room temperature in the supply storage room. 3. The laboratory reports performing 199,896 tests annually.</p>