

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D1025651	(X3) Date Survey Completed 06/11/2024
Name of Provider or Supplier Dermatology Center Of Canyon County	Street Address, City, State 318 2nd St S, Nampa, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on a direct observation and an interview with testing personnel 1 (TP1) on 6/11/2024, the laboratory failed to discontinue the use of expired potassium hydroxide (KOH). The findings include: 1. During the laboratory tour on 6/11/2024 a direct observation identified that the laboratory failed to discontinue the use of expired 20% KOH located in 11 of 11 patient rooms used in KOH slide examinations. EDM3 20% KOH lot 2122 expiration 5-2-2024 2. An interview with TP1 on 6/11/2024 at 2:45 pm confirmed the above finding. 3. The laboratory reports performing 300 KOH slide examinations annually.</p>