

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D1063459	(X3) Date Survey Completed 04/24/2025
Name of Provider or Supplier Idaho Falls Dermatology	Street Address, City, State 2001 S Woodruff Ave Ste 12a, Idaho Falls, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on a direct observation and an interview with the office manager on 4/24/2025, the laboratory failed to discontinue the use of expired urine Human Chorionic Gonadotropin (hCG) tests in 2024. The findings include: 1. A direct observation of the laboratory's hCG testing supplies on 4/24/2025 identified that the laboratory failed to discontinue the use of 37 Henry Schein Onestep+ urine hCG tests, lot HCG2052018, expiration 4/30/2024 prior to the expiration. 2. An interview with the office manager on 4/24/2025 at 2:07 pm confirmed the above finding. 3. The laboratory reports performing 150 waived tests annually.</p>