

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  13D1084633	<b>(X3) Date Survey Completed</b>  11/15/2022
<b>Name of Provider or Supplier</b>  Idaho Skin Surgery Center	<b>Street Address, City, State</b>  1906 S Vista Ave, Boise, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5217</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a review of potassium hydroxide (KOH) slide preparation documents and an interview with the laboratory lead on 11/15/2022, the laboratory failed to document, at least twice annually, the verification of accuracy for KOH slide preparations in 2021 . The findings include: 1. A review of KOH slide preparation bi-annual verification documentation identified that the laboratory failed to document verification of accuracy twice annually in 2021. 2. An interview with the laboratory lead on 11/15 /2022 at 9:38 am confirmed that the laboratory had not performed verification of accuracy for KOH slide preparations twice in 2021. 3. The laboratory reports performing 30 KOH slide preparations annually. 4. This is a repeat deficiency for KOH slide preparation bi-annual verification (5/30/2017).</p>
<b>D6120</b>	<p>TECHNICAL SUPERVISOR RESPONSIBILITIES CFR(s): 493.1451(b)(7)(8)</p> <p>(7) The technical supervisor is responsible for identifying training needs and assuring that each individual performing tests receives regular in-service training and education appropriate for the type and complexity of the laboratory services performed; (8) Evaluating the competency of all testing personnel and assuring that the staff maintain their competency to perform test procedures and report test results promptly, accurately and proficiently.</p> <p>This STANDARD is not met as evidenced by:</p>

Based on a review of the Centers for Medicare and Medicaid Services (CMS) 209 personnel form, training and competency assessment records and an interview with the laboratory lead on 11/15/2022, the technical supervisor failed to document initial training in 2022. The findings include: 1. A review of the CMS 209 identified one (1) new testing personnel performing moderate complexity testing hired after the last inspection(5/24/2021). 2. A review of training and competency assessment records identified one (1) of one (1) new testing personnel performing potassium hydroxide (KOH) slide preparations that failed to have documentation of initial training. 3. An interview with the laboratory lead on 11/15/2022 at 11:04 am confirmed there was no documentation of initial training for one (1) new testing personnel with a start date of February 2022. 4. The laboratory reports performing 30 KOH slide preparations annually.