Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D1096966	(X3) Date Survey Completed 07/22/2020
Name of Provider or Supplier Physicians Immediate Care Center	**Street Address, City, State** 134 W Chubbuck Rd, Chubbuck, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality. This STANDARD is not met as evidenced by: This Standard is not met as evidenced by: Based on observation of laboratory supplies and an interview with the lab manager, the laboratory failed to ensure expired vacutainer serum separator tubes were not available for use. Findings: 1. Observation of laboratory supplies available for use on July 22, 2020 at 1545 hours revealed that approximately 30 Becton Dickinson vacutainer serum separator tubes had expired in October 2019. 2. The laboratory manager confirmed on July 22, 2020 at 1600 hours that expired serum separator tubes were available for use in the laboratory.