

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D1096966	(X3) Date Survey Completed 01/10/2022
Name of Provider or Supplier Physicians Immediate Care Center	Street Address, City, State 134 W Chubbuck Rd, Chubbuck, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5211	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a record review of proficiency testing (PT) from the Wisconsin State Laboratory of Hygiene (WSLH) and an interview with the technical consultant (TC) on 1/10/2022, the laboratory failed to review and evaluate PT results that were less than 100% but were equal to or greater than 80%. The findings include: 1. A review of PT records from WSLH for 2020 event three identified that the laboratory failed to evaluate results for the analyte hematocrit which had a score of 80%. 2. An interview with the TC on 1/10/2022 at 1:38 pm confirmed that the laboratory failed to evaluate the PT hematocrit results for 2020 event three. 3. The laboratory reports performing 4500 hematology tests annually.</p>