

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  13D2000081	<b>(X3) Date Survey Completed</b>  03/08/2021
<b>Name of Provider or Supplier</b>  Idaho Skin Institute Of Chubbuck	<b>Street Address, City, State</b>  147 W Chubbuck Rd, Chubbuck, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D3031</b>	<p><b>RETENTION REQUIREMENTS</b> CFR(s): 493.1105(a)(3)</p> <p>Analytic systems records. Retain quality control and patient test records (including instrument printouts, if applicable) and records documenting all analytic systems activities specified in 493.1252 through 493.1289 for at least 2 years.</p> <p>This STANDARD is not met as evidenced by: Based on a record review and an interview with the laboratory technician on 3/8/2021, the laboratory failed to retain quality control (QC) records, temperature logs and maintenance logs for at least two (2) years. Then findings include: 1. A review of laboratory records for 2019 identified that the laboratory failed to retain room temperature and humidity logs, microscope maintenance logs, cryostat temperature logs, daily maintenance logs and QC staining logs for Hematoxylin and Eosin stains for two (2) years. 2. An interview with the laboratory technician on 3/8/2021 at 2:34 pm confirmed that the laboratory failed to retain temperature, maintenance and QC logs from 2019. 3. The laboratory reports performing 2000 tests annually.</p>