

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D2015997	(X3) Date Survey Completed 09/24/2019
Name of Provider or Supplier Beacon Pain Clinic	Street Address, City, State 115 W Main St Ste 201, Boise, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5439	<p>CALIBRATION AND CALIBRATION VERIFICATION CFR(s): 493.1255(b)</p> <p>Unless otherwise specified in this subpart, for each applicable test system the laboratory must do the following: Perform and document calibration verification procedure - (b)(1) Following the manufacturer's calibration verification instructions; (b)(2) Using the criteria verified or established by the laboratory under 493.1253(b)(3) -- (b)(2)(i) Including the number, type, and concentration of the materials, as well as acceptable limits for calibration verification; and (b)(2)(ii) Including at least a minimal (or zero) value, a mid-point value, and a maximum value near the upper limit of the range to verify the laboratory's reportable range of test results for the test system; and (b)(3) At least once every 6 months and whenever any of the following occur: (b)(3)(i) A complete change of reagents for a procedure is introduced, unless the laboratory can demonstrate that changing reagent lot numbers does not affect the range used to report patient test results, and control values are not adversely affected by reagent lot number changes. (b)(3)(ii) There is major preventive maintenance or replacement of critical parts that may influence test performance. (b)(3)(iii) Control materials reflect an unusual trend or shift, or are outside of the laboratory's acceptable limits, and other means of assessing and correcting unacceptable control values fail to identify and correct the problem. (b)(3)(iv) The laboratory's established schedule for verifying the reportable range for patient test results requires more frequent calibration verification.</p> <p>This STANDARD is not met as evidenced by: Based on a record review and an interview with the technical supervisor, the laboratory failed to perform and document calibration verification procedures at least once every 6 months for urine ethyl alcohol analyte performed on the Mindray BS-200 toxicology analyzer used to test drugs of abuse since the last survey on February 23, 2018. Findings: 1. A record review of calibration verification reports for urine ethyl alcohol revealed the laboratory failed to perform calibration verification</p>

activities since the last survey. 2. The laboratory performed approximately 15,000 urine ethyl alcohol the past year. 3. An interview on September 24, 2019, at 12:30 PM, with the technical supervisor, confirmed the laboratory failed to perform calibration verification activities on urine ethyl alcohol.