

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  13D2016348	<b>(X3) Date Survey Completed</b>  03/21/2024
<b>Name of Provider or Supplier</b>  Idaho Pain Clinic	<b>Street Address, City, State</b>  1327 Superior St #101, Sandpoint, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5217</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a review of proficiency testing (PT) records from College of American Pathologists (CAP) and an interview with the technical supervisor (TS) on 3/21/2024, the laboratory failed to enroll in proficiency testing for all tested analytes in 2022 and 2023. The findings include: 1. A review of proficiency testing records from CAP identified that the laboratory failed to enroll and participate in PT that covered all analytes performed by LCMSMS at least two times annually in 2022 and 2023. The analytes tested include: methadone amphetamine, methamphetamine, alprazolam, aOH- alprazolam, Clonazepam, 7-aminoclonazepam, diazepam, temazepam, oxazepam, nordiazepam, lorazepam, Fentanyl, EDDP, meperidine, normeperidine, Tramadol, O-Tramadol, phencyclidine, MDMA, MDA, MDEA, MDPV, 6-acetylmorphine, benzoylecgonine, mephedrone, ehtylone, methylone, codeine, morphine, hydromorphone, hydrocodone, dihydrocodeine, oxycodone, oxymorphone, buprenorphine, norbupreneophine, naloxone, carisoprodol, meprobamate, zolpidem, nortriptyline, Ketamine norketamine, Gabapentin, tapentadol, cotinine. 2. An interview with the TS on 3/21/2024 at 1:36 pm confirmed that the laboratory failed to enroll and participate in PT for all analytes performed by LCMSMS at least twice annually for 2022 and 2023. 3. The laboratory reports performing 249,600 LCMSMS toxicology tests annually.</p>