

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D2167724	(X3) Date Survey Completed 05/26/2026
Name of Provider or Supplier Mountain Pine Dermatology PLLC	Street Address, City, State 211 E Logan St Ste 305, Caldwell, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on a direct observation and an interview with the laboratory lead on 5/26/2026, the laboratory failed to discontinue the use of expired Mohs reagents since April 2026. The findings include: 1. A direct observation of the laboratory's reagents on 5/26/2026 identified that the laboratory failed to discontinue the use of the Eosin-Y Vintage stain lot 208517 expiration 2026-03-31 prior to the expiration. 2. An interview with the laboratory lead on 5/26/2026 at 11:12 am confirmed the above finding. 3. The laboratory reports performing 275 Mohs slide examinations annually.</p>