

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D2173071	(X3) Date Survey Completed 08/11/2026
Name of Provider or Supplier Idaho Falls Community Hospital	Street Address, City, State 2327 Coronado St, Idaho Falls, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235 As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency. This STANDARD is not met as evidenced by: Based on a review of the Centers for Medicare and Medicaid Services (CMS) 209 personnel form, laboratory procedures, training and competency assessment records and an interview with the technical consultant, the laboratory failed to follow written procedures to assess testing personnel competency in 2025 and 2026. The findings include: 1. The CMS 209 identified 151 testing personnel (TP) performing moderate complexity testing. 2. A review of laboratory procedures identified that the laboratory established a procedure to assess TP initial training, semiannual and annual competency. 3. A review of training and competency assessment records identified that the laboratory failed to have six (6) month competency assessments for seven (7) TP in 2025. 4. A review of training and competency assessment records identified that the laboratory failed to have annual competency assessments for five (5) TP in 2025 and three (3) TP in 2026. 5. An interview with the technical consultant on 8/11/2026 at 11:39 am confirmed the above findings. 6. The laboratory reports performing 139,200 moderate complexity tests annually. 7. This is a repeat deficiency from previous inspections on 11/6/2020, 7/11/2022 and 9/10/2024. The laboratory failed to implement and follow their plans of correction.