Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D2231041	(X3) Date Survey Completed 07/13/2023
Name of Provider or Supplier Beacon Cancer Care, Pllc	**Street Address, City, State** 3815 N Schreiber Way Ste 101, Coeur D Alene, ID	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2007	TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(1) The samples must be examined or tested with the laboratory's regular patient workload by personnel who routinely perform the testing in the laboratory, using the laboratory's routine methods This STANDARD is not met as evidenced by: Based on a review of proficiency testing (PT) records from the American Proficiency Institute (API), the Centers for Medicare and Medicaid Services (CMS) 209 personnel form, competency assessments and an interview with the laboratory supervisor on 7/13/2023, the laboratory failed to test PT samples with the regular patient workload by personnel who regularly perform the laboratory testing. The findings include: 1. A review of the CMS 209 and competency assessments identified four (4) testing personnel performing complete blood count (CBC) testing on the Sysmex XN430 analyzer. 2. A review of PT records from API for hematology for 2022 events one, two and three and 2023 event one identified that one (1) of four (4) testing personnel performed all PT events for 2022 and 2023. The laboratory failed to have three (3) of four (4) personnel who regularly perform CBC testing perform PT testing in 2022 and 2023. 3. An interview with the laboratory supervisor on 7/13/2023 at 2:10 pm confirmed that one (1) of four (4) testing personnel performed all the CBC PT testing for 2022 and 2023. 4. The laboratory reports performing 12,000 CBC tests annually.