

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 14D0424105	(X3) Date Survey Completed 04/19/2023
Name of Provider or Supplier Rush Univ Med Center Dermatology Patient Svcs	Street Address, City, State 1725 W Harrison St, Ste 264, Chicago, IL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory records, lack of documentation, and interview with the laboratory consultant (LC), the laboratory failed to evaluate the biannual method accuracy for potassium hydroxide (KOH) mycology testing at least twice a year in 2021 and 2022. Findings include: 1. Review of the "Competency and Proficiency Testing Policy" revealed: a. "For Provider Performed Microscopy (PPM) testing: The Dermatology clinic utilizes the API program for KOH and CAP program for Scabies." 2. Review of laboratory proficiency testing records and lack of documentation revealed that the laboratory failed to evaluate and document the biannual method accuracy for potassium hydroxide (KOH) testing in 2021 and 2022. Affecting a total of 31 test results: a. Annual KOH testing volume (2021) = 7 b. Annual KOH testing volume (2022) = 24 3. On 04/19/2023 at 11:15 a.m., LC #2 stated, "The laboratory failed to enroll in proficiency testing for KOH in 2021 and 2022".</p>