

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 14D0432728	(X3) Date Survey Completed 06/03/2026
Name of Provider or Supplier Champaign Urbana Public Health District	Street Address, City, State 201 West Kenyon Rd, Champaign, IL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6054	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9)</p> <p>(b)(9) Thereafter, evaluations must be performed at least annually</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's policy and procedure manual, testing personnel competency records, and interview with Technical Consultant (TC) #1; the laboratory failed to have competency assessments performed annually for one of five testing personnel (TP) performing moderate complexity testing for the specialty of Microbiology. Findings include: 1. Review of the laboratory policy titled, "Laboratory Personnel Competency testing", stated the following, "All CUPHD testing personnel (TP) will be observed semi-annually in the first year of testing patient specimens and annually thereafter. The technical consultant (TC) will be responsible for assessing competency." 2. Review of testing personnel competency records revealed one of five TP (TP #4) performing moderate complexity testing failed to have competency evaluations completed annually in 2024. 3. Interview with TC #1 on 06/03/2026, at 11:06 am, confirmed the laboratory failed to have completed competency assessments annually for one of five TP performing the moderate complexity testing.</p>