

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 14D0433367	(X3) Date Survey Completed 10/04/2021
Name of Provider or Supplier Thomas H Boyd Memorial Hospital	Street Address, City, State 800 School Street, Carrollton, IL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on review of CASPER Report 0155D and interview with an American Proficiency Institute (API) representative the laboratory failed to successfully participate in proficiency testing (PT) for the immunohematology analyte compatibility testing during events 1 and 2 of 2021. See D2181.</p>
D2181	<p>COMPATIBILITY TESTING CFR(s): 493.863(e)</p> <p>Failure to achieve an overall testing event score of satisfactory for two consecutive</p>

testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on review of CASPER Report 0155D and interview with an American Proficiency Institute (API) representative the laboratory failed to successfully participate in proficiency testing (PT) for compatibility testing in events 1 and 2 of 2021, under the specialty of immunohematology. Findings include: 1. Review of the CASPER Report 0155D revealed that the initial unsuccessful PT performance occurred during API PT events 1 and 2 of 2021, as listed below.

IMMUNOHEMATOLOGY EVENT -1, 2021 Compatibility Testing, 80%

Unsatisfactory EVENT -2, 2021 Compatibility Testing, 80% Unsatisfactory 2.

Interview with the API PT representative on 10-04-2021, at 9:32 AM confirmed the unsuccessful PT performance for compatibility testing in events 1 and 2 of 2021.