

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 14D0678742	<b>(X3) Date Survey Completed</b> 12/16/2025
<b>Name of Provider or Supplier</b> Pinski Dermatology & Cosmetic Surgery	<b>Street Address, City, State</b> 150 N Michigan Ave, Ste 1200, Chicago, IL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5217</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory records, lack of documentation, and interview with laboratory representative; the laboratory failed to perform biannual method accuracy (proficiency testing/peer reviewed histopathology interpretations) for dermatopathology for two of two years reviewed, 2024 and 2025. Findings include: 1. Review of laboratory records revealed the laboratory sent slides out for peer review one time each year for dermatopathology interpretations in 2024 and 2025. Date(s) of testing: Date of review: 01/12/2024 - 12/16/2024 01/29/2025 01/27/2025 - 10/01/2025 11/05/2025 2. Interview with the laboratory representative on 12/16/2025, at 1:57 pm, confirmed the laboratory failed to perform biannual method accuracy (proficiency testing/peer reviewed histopathology interpretations) for dermatopathology for two of two years reviewed, 2024 and 2025.</p>
<b>D5221</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(d)</p> <p>All proficiency testing evaluation and verification activities must be documented.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory records, lack of documentation, and interview with the laboratory representative; the laboratory failed to evaluate/review results of biannual method accuracy (proficiency testing/peer reviewed histopathology interpretations) for two of two dermatopathology events performed in 2024 and 2025. Findings</p>

include: 1. Review of laboratory records revealed the laboratory failed to evaluate peer reviewed dermatopathology interpretations in 2024 and 2025. Date(s) of testing: Date of review: 01/12/2024 - 12/16/2024 01/29/2025 01/27/2025 - 10/01/2025 11/05/2025 2. Interview with the laboratory representative on 12/16/2025, at 1:57 pm, confirmed the laboratory failed to evaluate/review results of biannual method accuracy (proficiency testing/peer reviewed histopathology interpretations) for the above two dermatopathology peer reviews in 2024 and 2025.