

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 14D0697391	(X3) Date Survey Completed 10/04/2023
Name of Provider or Supplier Tony Fu Dermatology	Street Address, City, State 636 Church St #308, Evanston, IL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory records, lack of documentation and interview with the laboratory director (LD), the laboratory failed to make available a written procedures manual for one of one high complexity histopathology microscopic slide evaluation and interpretation with diagnostic reporting testing procedure in 2021, 2022 and 2023. Findings include: 1. Surveyor review of laboratory records and lack of documentation revealed the laboratory failed to provide laboratory personnel a written procedures manual for high complexity histopathology microscopic slide evaluation and interpretation with diagnostic reporting testing procedure in 2021, 2022 and 2023. 2. On 10/04/2023 at 09:14 a.m., the LD confirmed the above findings.</p>