

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  14D0697391	<b>(X3) Date Survey Completed</b>  10/22/2025
<b>Name of Provider or Supplier</b>  Tony Fu Dermatology	<b>Street Address, City, State</b>  636 Church St #308, Evanston, IL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5217</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory records, lack of documentation, and interview with laboratory director (LD); the laboratory failed to perform bi-annual evaluations for one of one high complexity test system: dermatopathology testing in 2024. Findings Include: 1. Review of laboratory records found no bi-annual method accuracy evaluations for dermatopathology testing in 2024. 2. Interview with the LD on 10-22-25 at 10:09 am confirmed that no bi-annual method accuracy evaluation had been performed for Dermatopathology testing in 2024.</p>
<b>D5221</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(d)</p> <p>All proficiency testing evaluation and verification activities must be documented.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory records, lack of documentation, and interview with the laboratory director (LD), the laboratory failed to evaluate results of bi-annual method accuracy (proficiency testing/peer reviewed histopathology interpretations) for two of two events from 2025. Findings include: 1. Review of laboratory records revealed no documented evaluations of results upon receipt of peer reviewed histopathology interpretations for two of two bi-annual method accuracy events in 2025. 2. Interview with the LD on 10/22/2025, at 10:10 am, confirmed the laboratory failed to document evaluation of results of peer reviewed histopathology interpretations in 2025.</p>