

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 14D0969917	(X3) Date Survey Completed 05/26/2022
Name of Provider or Supplier Fertility Centers Of Illinois	Street Address, City, State 907 N Elm St, Hinsdale, IL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5807	<p>TEST REPORT CFR(s): 493.1291(d)</p> <p>Pertinent "reference intervals" or "normal" values, as determined by the laboratory performing the tests, must be available to the authorized person who ordered the tests and, if applicable, the individual responsible for using the test results.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory records and interview with the laboratory technical supervisor (TS); the laboratory failed to report complete Viscosity Reference Ranges for semen analysis on four of four final patient reports. Findings Include: 1. Laboratory procedures manual and final reports of four patients (PT) from dates; 10/09/2020 (PT 1), 11/22/2021 (PT 2), 10/01/2021 (PT 3), and 01/04/2022 (PT 4) were reviewed. 2. Surveyor review of four of four patient (PT) final reports revealed: PATIENT VISC. RES* RR** (PT 1) Very increased Average (PT 2) Average Average (PT 3) Average Average (PT 4) Average Average 3. Review of the reference ranges chart utilized for examining patient samples on page 32 of the laboratory procedure manual revealed: Result RR** Viscosity Average/Increased/Decreased 4. On survey date 05/26/2022, at 10:50 am, the above findings were confirmed by the TS. * VISCOSITY RESULT ** REFERENCE RANGE (WHO 5TH EDITION)</p>