

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 14D2116100	(X3) Date Survey Completed 09/24/2025
Name of Provider or Supplier Beaird Dermatology S C	Street Address, City, State 4885 Hoffman Blvd Ste 407, Hoffman Estates, IL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: D5217 Based on review of laboratory policies and procedures, laboratory proficiency testing records, lack of documentation, and interview with the technical supervisor (TS), the laboratory failed to perform bi-annual method accuracy evaluations for microscopic Potassium Hydroxide (KOH) wet mount testing from 2023 to the date of survey, 09/24/2025. Findings include: 1. Review of laboratory policies and procedures revealed the policy titled, "General Laboratory Policies", which stated, under section titled "7.Proficiency Testing", "This laboratory will enroll in formal proficiency testing appropriate to the test menu and treat all proficiency test challenge specimens the same as patient specimens." 2. Review of laboratory PT records found no documented KOH proficiency testing or bi-annual method accuracy evaluations. 3. Interview with the TS on 09/24/2025, at 9:48 am, confirmed that the laboratory was not enrolled in proficiency testing nor did the laboratory have any documented bi-annual method accuracy evaluations for KOH testing.</p>
D5291	<p>GENERAL LABORATORY SYSTEMS QUALITY ASSESSMENT CFR(s): 493.1239(a)</p> <p>The laboratory must establish and follow written policies and procedures for an ongoing mechanism to monitor, assess, and, when indicated, correct problems identified in the general laboratory systems requirements specified at 493.1231 through 493.1236.</p>

This STANDARD is not met as evidenced by:

A Based on review of the laboratory's quality assessment policies and procedures, review of quality assurance records, review of proficiency testing (PT) records, and interview with the technical supervisor (TS), the laboratory failed to follow the written quality assurance policy for monitoring, assessing, and correcting problems related to PT from 2023 to the date of survey, 09/24/25. Findings Include: 1. Review of the laboratory's policy and procedure manual identified the policy, "Quality Assurance Program", stated the following: "Twice yearly review by the laboratory director or designee will be conducted for quality assurance. This will cover the quality assessment program for procedures used in this laboratory. The checklists are used to evaluate general laboratory systems, Pre-analytical systems, analytic systems, and post-analytic systems. Any discrepancy found in the checklist will be documented as a lab error form and kept I the manual for two years ... Error reporting Anytime and every time an error occurs, an error form must be filled out." 2.Review documents titled "Quality assurance (QA) annual review" signed by the laboratory director /designee on 3/21/25 and 3/15/24 did not identify issues related to PT and did not have documented corrective actions. i.Review of PT testing records found no bi-annual method accuracy evaluations for KOH preparations. See D5217. ii.Review of PT records found that bi-annual PT for Histopathology in 2023 was not completed as defined by the sites policy and procedures, 2023 PT review documented completion on 01/03/25 without a documented corrective action. 3. Interview with the TS on 09-24-2025 at 9:25 am confirmed the laboratory had failed to follow the quality assessment policies as described in the document "Quality Assurance Program". B Based on review of the laboratory's quality assessment policies and procedures, review of laboratory temperature and humidity logs, review of quality assurance records and interview with the technical supervisor (TS), the laboratory failed to follow the written quality assurance policy for monitoring, assessing, and correcting problems related to temperature monitoring for 2023 and 2024 affecting 210 of 488 days the laboratory was performing testing. 1. Review of laboratory policies and procedures revealed the policy titled, "General Laboratory Policies", which stated, under the section titled "12.Documentation of QA and QC", "We document the results of all QA checks and controls on log sheets and will take appropriate actions when equipment checks fall outside of the established parameters or controls do not perform as expected." 2. Review of Laboratory temperature and humidity monitoring logs found 210 of 488 dates to have been outside of the laboratories acceptable range for temperature or humidity. See D5781. 3. Review of documents titled "QA monthly checklist" dated January 2024 and completed by TP #3 and reviewed by the LD/ Designee did not identify the issues related to temperature and humidity monitoring and did not have documented corrective actions. 4. Interview with the TS on 09-24-2025 at 11:32 am confirmed the laboratory had failed to follow the quality assessment policies as described in the document "General Laboratory Policies".

D5781

CORRECTIVE ACTIONS
CFR(s): 493.1282(b)(1)

(b) The laboratory must document all corrective actions taken, including actions taken when any of the following occur: (b)(1) Test systems do not meet the laboratory's verified or established performance specifications, as determined in 493.1253(b), which include but are not limited to-- (b)(1)(i) Equipment or methodologies that perform outside of established operating parameters or performance specifications; (b)(1)(ii) Patient test values that are outside of the laboratory's reportable range of test results for the test system; and (b)(1)(iii) When the laboratory determines that the reference intervals (normal values) for a test procedure are inappropriate for the

laboratory's patient population.

This STANDARD is not met as evidenced by:

Based on review of laboratory records and interview with the office technical supervisor (TS); the laboratory failed to perform corrective actions for out of range temperature/ humidity readings for 210 of 488 dates when histopathology testing was performed. 1. The laboratory's 2023 and 2024 "Room Temperature & Humidity Log" indicate temperature should be 20-28 C[Celsius] (68-82 F[Fahrenheit]) and humidity should be "30-50% ". 2. Review of the 2023 and 2024 "Room Temperature & Humidity Log" records identified that the temperature/humidity was outside than the laboratory's established acceptability criteria for 210 of 488 histopathology testing dates reviewed. Testing Date Temperature/Humidity (%) 1. 01-05-23 72.8 F/27.1 2. 01-09-23 72.5 F/25.1 3. 01-10-23 73 F/26.6 4. 01-11-23 73.7 F/28.8 5. 01-12-23 74.2 F/29.5 6. 01-13-23 75.1 F/29.7 7. 01-16-23 71.1 F/23.1 8. 01-18-23 72.2 F/28.0 9. 01-19-23 72.5 F/29.1 10. 01-23-23 71.0 F/24.4 11. 01-24-23 71.7 F/24.4 12. 01-25-23 73.4 F/25.1 13. 01-26-23 73.6 F/25.5 14. 01-30-23 70.0 F/18.4 15. 01-31-23 69.7 F/15.1 16. 02-01-23 71.1 F/13.9 17. 02-02-23 70.8 F/75.8 18. 02-03-23 72.0 F/73.6 19. 02-06-23 71.3 F/20.1 20. 02-07-23 72.5 F/25.2 21. 02-08-23 72.7 F/24.9 22. 02-09-23 72.4 F/28.2 23. 02-13-23 72.0 F/23.0 24. 02-14-23 72.1 F/22.3 25. 02-16-23 71.4 F/25.3 26. 02-18-23 70.0 F/21.0 27. 02-20-23 71.6 F/24.9 28. 02-21-23 72.9 F/21.9 29. 02-22-23 73.4 F/22.7 30. 02-23-23 72.5 F/26.4 31. 02-24-23 73.3 F/22.4 32. 02-27-23 70.9 F/28.7 33. 02-28-23 71.1 F/29.7 34. 03-01-23 72.5 F/29.5 35. 03-02-23 72.3 F/27.7 36. 03-03-23 72.5 F/26.3 37. 03-07-23 73.2 F/29.6 38. 03-08-23 72.2 F/21.8 39. 03-09-23 71.7 F/19.7 40. 03-10-23 72.9 F/24.8 41. 03-13-23 70.2 F/26.2 42. 03-14-23 70.5 F/20.2 43. 03-15-23 71.1 F/20.0 44. 03-16-23 72.9 F/23.3 45. 03-17-23 72.5 F/25.3 46. 03-21-23 73.0 F/19.9 47. 03-22-23 72.0 F/28.1 48. 03-24-23 73.7 F/26.2 49. 03-27-23 71.0 F/29.3 50. 03-28-23 70.8 F/23.3 51. 03-29-23 22.0 C/25.9 52. 03-30-23 21.6 C/18.8 53. 03-31-23 22.9 C/29.9 54. 04-03-23 21.9 C/26.3 55. 04-06-23 22.5 C/29.5 56. 04-10-23 22.6 C/24.8 57. 04-18-23 22.6 C/25.9 58. 04-19-23 22.6 C/25.8 59. 04-24-23 21.7 C/26.4 60. 04-25-23 22.2 C/26.9 61. 04-26-23 22.3 C/26.0 62. 04-27-23 22.5 C/24.5 63. 04-28-23 23.0 C/27.4 64. 05-02-23 22.6 C/28.0 65. 05-03-23 22.7 C/26.3 66. 05-04-23 21.7 C/29.6 67. 05-18-23 23.9 C/28.6 68. 05-25-23 73.1 F/29.3 69. 06-30-23 74.8 F/50.8 70. 07-06-23 72.3 F/54.3 71. 07-20-23 73.5 F/52.8 72. 07-28-23 74.7 F/51.4 73. 07-29-23 72.6 F/51.1 74. 08-01-23 73.4 F/51.3 75. 08-02-23 73.4 F/51.3 76. 08-05-23 73.4 F/59.7 77. 08-11-23 71.9 F/52.1 78. 08-15-23 72.2 F/51.1 79. 08-17-23 72.3 F/51.5 80. 08-23-23 75.2 F/55.5 81. 08-24-23 75.6 F/56.4 82. 08-25-23 74.0 F/50.5 83. 08-29-23 72.5 F/50.7 84. 09-03-23 83.3 F/45.5 85. 09-06-23 73.0 F/51.7 86. 09-07-23 73.0 F/50.8 87. 09-19-23 73.1 F/51.8 88. 09-21-23 71.6 F/53.2 89. 09-22-23 73.4 F/51.5 90. 09-26-23 73.3 F/55.2 91. 09-27-23 73.8 F/53.1 92. 09-28-23 72.9 F/54.1 93. 09-29-23 73.7 F/53.0 94. 10-10-23 72.3 F/27.1 95. 10-11-23 72.5 F/26.2 96. 10-26-23 72.5 F/51.9 97. 10-27-23 74.4 F/53.7 98. 11-13-23 73.0 F/27.0 99. 11-14-23 22.6 C/22.4 100. 11-15-23 22.7 C/22.7 101. 11-16-23 22.3 C/28.8 102. 11-20-23 21.9 C/28.5 103. 11-22-23 24.5 C/29.6 104. 11-27-23 20.7 C/21.0 105. 11-28-23 21.3 C/17.5 106. 11-29-23 21.5 C/17.5 107. 11-30-23 22.4 C/22.3 108. 12-05-23 21.4 C/29.0 109. 12-06-23 21.6 C/26.5 110. 12-07-23 21.7 C/26.7 111. 12-10-23 20.8 C/27.6 112. 12-11-23 21.8 C/25.5 113. 12-12-23 21.8 C/22.2 114. 12-13-23 22.7 C/20.8 115. 12-14-23 22.4 C/25.3 116. 12-18-23 21.2 C/28.0 117. 12-19-23 20.7 C/21.2 118. 12-20-23 21.1 C/22.0 119. 12-21-23 21.6 C/28.7 120. 12-28-23 22.3 C/29.5 121. 01-02-24 20.9 C/24.7 122. 01-03-24 21.4 C/26.0 123. 01-04-24 21.6 C/23.6 124. 01-05-24 21.5 C/22.8 125. 01-08-24 21.2 C/26.0 126. 01-09-24 21.3 C/27.8 127. 01-10-24 21.5 C/27.6 128. 01-11-24 21.5 C/27.2 129. 01-12-24 21.8 C/27.3 130. 01-15-24 18.6 C/16.2 131. 01-16-24 19.3 C/19.4 132. 01-17-24 20.1 C/19.5 133. 01-18-24 20.2 C/17.2

1 134. 01-19-24 21.0 C/18.2 135. 01-22-24 20.7 C/17.9 136. 01-23-24 22.4 C/24.0
 137. 01-24-24 22.5 C/27.6 138. 01-29-24 19.9 C/31.7 139. 02-01-24 21.8 C/29.3 140.
 02-05-24 21.3 C/28.3 141. 02-06-24 22.2 C/28.0 142. 02-07-24 22.4 C/28.0 143. 02-
 08-24 21.9 C/28.8 144. 02-10-24 22.9 C/26.3 145. 02-12-24 20.8 C/25.9 146. 02-13-
 24 21.5 C/24.1 147. 02-14-24 21.4 C/21.9 148. 02-15-24 21.8 C/27.3 149. 02-16-24
 22.0 C/20.1 150. 02-19-24 21.3 C/20.4 151. 02-20-24 22.1 C/21.7 152. 02-21-24 21.9
 C/26.4 153. 02-22-24 21.9 C/28.6 154. 02-23-24 23.4 C/25.6 155. 02-26-24 21.3 C/25.
 6 156. 03-07-24 22.3 C/26.5 157. 03-11-24 20.8 C/25.6 158. 03-18-24 21.3 C/24.1
 159. 03-19-24 22.0 C/21.2 160. 03-21-24 22.0 C/18.2 161. 03-22-24 22.9 C/21.3 162.
 03-27-24 71.4 F/24.8 163. 03-28-24 71.7 F/24.4 164. 03-29-24 70.5 F/23.3 165. 04-05-
 24 75.7 F/26.1 166. 04-19-24 72.3 F/25.6 167. 04-22-24 73.4 F/21.8 168. 04-24-24
 72.7 F/25.4 169. 04-25-24 73.4 F/22.8 170. 06-03-24 73.8 F/51.9 171. 06-04-24 74.3 F
 /52.9 172. 06-05-24 73.0 F/52.1 173. 06-18-24 73.8 F/50.1 174. 06-20-24 73.3 F/50.9
 175. 06-21-24 73.1 F/53.1 176. 06-26-24 73.8 F/57.2 177. 07-03-24 73.3 F/53.5 178.
 07-05-24 74.6 F/52.0 179. 07-10-24 73.1 F/57.4 180. 07-15-24 74.4 F/50.9 181. 07-16-
 24 73.4 F/54.7 182. 07-17-24 73.2 F/54.3 183. 07-23-24 73.0 F/50.3 184. 07-24-24
 73.8 F/50.5 185. 07-26-24 74.3 F/51.3 186. 07-30-24 74.1 F/54.6 187. 07-31-24 74.1 F
 /52.1 188. 08-01-24 73.5 F/53.0 189. 08-02-24 74.3 F/54.6 190. 08-08-24 73.8 F/50.6
 191. 08-13-24 73.2 F/51.6 192. 08-14-24 73.6 F/51.2 193. 08-15-24 73.1 F/54.2 194.
 08-16-24 73.4 F/53.0 195. 08-20-24 73.7 F/50.6 196. 08-27-24 74.7 F/51.7 197. 08-29-
 24 73.2 F/51.0 198. 08-30-24 73.9 F/52.6 199. 09-18-24 74.6 F/51.5 200. 09-25-24
 74.3 F/50.5 201. 10-16-24 74.3 F/28.6 202. 10-17-24 74.4 F/24.8 203. 10-18-24 75.2 F
 /27.6 204. 10-28-24 72.0 F/28.3 205. 12-05-24 72.8 F/19.5 206. 12-06-24 72.3 F/19.6
 207. 12-18-24 73.7 F/27.4 208. 12-19-24 73.4 F/23.7 209. 12-20-24 73.2 F/25.6 210.
 12-23-24 71.4 F/25.4 3. On survey date 09-24-2025, at 11:32 am, the TS confirmed
 no corrective actions were taken for out of range humidity/temperature readings when
 histopathology testing was performed.

D6093

LABORATORY DIRECTOR RESPONSIBILITIES
 CFR(s): 493.1445(e)(5)

(e)(5) Ensure that the quality control and quality assessment programs are established and maintained to assure the quality of laboratory services provided and to identify failures in quality as they occur;

This STANDARD is not met as evidenced by:

Based on review of laboratory policies and procedures, review of laboratory quality assurance records and interview with the technical supervisor (TS), the laboratory director failed to follow written quality assessment policies and procedures for monitoring, assessing, and correcting problems for the laboratory. See 5291 A & B