

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 14D2121950	(X3) Date Survey Completed 06/17/2024
Name of Provider or Supplier Dermatology Associates Of Lombard	Street Address, City, State 500 East 22nd St, Lombard, IL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation of laboratory testing supplies and interview with the laboratory representative; the laboratory failed to ensure five of five margin marking dyes used for the subspecialty of histopathology were not used when they had exceeded their expiration date. Findings Include: 1. Upon a tour of the laboratory on 06-17-2024, at 12:05 pm, direct observation of laboratory testing supplies found five of five margin marking dyes, used for demonstrating tissue margins in the grossing process of histopathology, to be expired. Margin Marker Dye: Expiration: Product #: Yellow, 2 oz 06-30-2022 SL662YL-2 Blue, 2 oz 05-31-2022 SL662BL-2 Red, 2 oz 05-31-2022 SL662RD-2 Black, 2 oz 01-31-2024 SL662BK-2 Green, 2 oz 10-31-2018 SL662GR-2 2. On survey date 06-17-2024, at 12:35 pm, the above findings were confirmed by the laboratory representative.</p>