

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 14D2153680	(X3) Date Survey Completed 12/03/2019
Name of Provider or Supplier Premier Urgent Care	Street Address, City, State 1301 E 47th St, Chicago, IL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on surveyor review of CASPER Report 155 Proficiency Testing (PT) records, and phone communication with the PT vendor College of American Pathologists (CAP); it was confirmed that the laboratory failed to successfully participate in the testing of PT samples under the specialty of Chemistry. Findings include: 1. Review of PT CASPER Report 155 on December 03, 2019 at 12:02 PM revealed that the unsatisfactory PT performance was a Subsequent -not the first, that occurred during calendar year 2019. See D2096.</p>
D2096	ROUTINE CHEMISTRY

CFR(s): 493.841(f)

Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on surveyor review of CASPER Report 155 Proficiency Testing (PT) records, and phone communication with the PT vendor College of American Pathologists (CAP); it was confirmed that the laboratory failed to successfully participate in the testing of PT samples under the specialty of Chemistry. Findings include: 1. Review of PT CASPER Report 155 on December 03, 2019 at 12:02 PM, revealed that the unsatisfactory PT performance was a subsequent -not the first, that occurred under the subspecialty of Routine Chemistry for the analyte Sodium (Na) during calendar year 2019, as listed below. CHEMISTRY Routine Chemistry EVENT-1, 2019 Sodium (Na) = 60% EVENT-2, 2019 Sodium (Na) = 40% EVENT-3, 2019 Sodium (Na) = 40% 2. During PT Desk review on December 03, 2019 and phone communication with the PT vendor CAP at 12:02 PM, the Subsequent PT failing scores were confirmed.