

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  15D2108281	<b>(X3) Date Survey Completed</b>  04/29/2026
<b>Name of Provider or Supplier</b>  Ascension Medical Group St Vincent - Dermatology	<b>Street Address, City, State</b>  3801 Bellemeade Ave, Suite 120, Evansville, IN	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5217</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on proficiency testing review, policy and procedure review, medical record review and interview, the laboratory failed to perform twice annual verification for KOH in 2025. Findings include: 1. Review of CAP proficiency testing showed no results for FSM-A 2025, KOH testing samples in 2025. 2. Policy titled, "KOH Proficiency testing", written on 04/01/2016 stated, "Twice annually The College of American Pathologist will send KOH proficiency kits to test the physician 's ability to diagnosis various fungal specimen types. The answer sheet will be faxed to CAP ... Results will be kept on file in the laboratory ..." 3. A KOH fungal sample was analyzed microscopically on the following patient: Patient Collection Date Reported Results P1 11/18/2025 Hyphae Present 4. In an interview on 04/29/2026 at 4:15 pm, E1 (Office Manager) confirmed the laboratory had completed twice annual verification only once for KOH testing in 2025. Legend: CAP-College of American Pathologist E-Employee FSM-Fungal Smear KOH-Potassium Hydroxide Preparation P-Patient</p>