

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 16D0038924	(X3) Date Survey Completed 06/18/2026
Name of Provider or Supplier Henry County Health Center Inc	Street Address, City, State 407 South White Street, Mount Pleasant, IA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5215	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2) The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results). This STANDARD is not met as evidenced by: Based on review of American Proficiency Institute (API) proficiency testing (PT) records and confirmed by interview with Technical Supervisor #1 (TS #1) at 10:05 am on 06/18/2026, the laboratory failed to perform and document a self evaluation when the laboratory received six ungraded PT scores from four out of seven PT testing events from 01/01/2024- 06/18/2026. The findings include: 1. For 2024 testing event 2, the laboratory received ungraded PT test scores for the following: *2024 Hematology/Coagulation- vaginal wet prep (specimen VA-02) *2024 Microbiology- gram stain (specimen GS-08) 2. For 2025 testing event 2, the laboratory received ungraded PT test scores for the following: *2025 Microbiology- gram stain morphology (specimen GS-08) 3. For 2025 testing event 3, the laboratory received ungraded PT test scores for the following: *2025 Hematology/Coagulation- blood cell identification (BCI-15) 4. For 2026 testing event 1, the laboratory received ungraded PT test scores for the following: *2026 Microbiology- gram stain morphology (specimen GS-04) and blood culture (specimen BLD-02) 5. At the time of the survey, TS #1 confirmed the laboratory failed to perform and document a self evaluation for the ungraded PT test scores listed above.
D5219	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(2)

(c)(2) Any test or procedure listed in subpart I of this part for which compatible proficiency testing samples are not offered by a CMS-approved proficiency testing program.

This STANDARD is not met as evidenced by:

Based on review of the Laboratory Test List & Annual Volume form, lack of verification of accuracy records, and confirmed by interview with Technical Supervisor #1 (TS #1) at 10:05 am on 06/18/2026, the laboratory failed to verify the accuracy of *Helicobacter pylori* rapid urea testing at least twice annually for two out of two time periods from 01/01/2025- 12/31/2025. The findings include: 1. The Laboratory Test List & Annual Volume form listed *Helicobacter pylori* rapid urea testing. 2. TS #1 confirmed the laboratory began using the Remel Selective Rapid Urea Kit for *Helicobacter pylori* rapid urea testing in November 2024. 3. At the time of the survey, TS #1 confirmed the laboratory failed to verify the accuracy of the *Helicobacter pylori* rapid urea testing at least twice annually from 01/01/2025- 12/31/2025.