Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 16D0382708	(X3) Date Survey Completed 01/20/2026
Name of Provider or Supplier South Central Iowa Medical Clinic	**Street Address, City, State** 417 South East Street Suite 100, Corydon, IA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5783	CORRECTIVE ACTIONS CFR(s): 493.1282(b)(2) (b)(2) Results of control or calibration materials, or both, fail to meet the laboratory's established criteria for acceptability. All patient test results obtained in the unacceptable test run and since the last acceptable test run must be evaluated to determine if patient test results have been adversely affected. The laboratory must take the corrective action necessary to ensure the reporting of accurate and reliable patient test results. This STANDARD is not met as evidenced by: Based on review of the DCA Vantage Individualized Quality Control Plan (IQCP), quality control (QC) instrument tapes, patient test reports and confirmed by interview with testing personnel identifier #1 (PT #1) at approximately 2:00 pm on 1/20/2026, the laboratory failed to take and document corrective action when albumin and creatinine QC results failed to meet the laboratory's established criteria for acceptability for two out of nine lot numbers of DCA Vantage cartridges performed between 1/1/2025 - 12/31/2025. The findings include: 1. The DCA Vantage IQCP stated the laboratory would perform external quality control with each lot number and /or shipment of DCA Vantage cartridges. 2. The DCA Vantage QC instrument tapes listed the acceptable range for the high level of QC for lot numbers 9134 and 0190 as: *Albumin 168.0 - 252 mg/L *Creatinine 352.6 - 488.4 mg/dL 3. On 2/28/2025 the laboratory performed QC on lot number 9134 of DCA Vantage cartridges and received an out of range QC results of 75.4 mg/L for albumin and 151.7 mg/dL for creatinine. 4. On 7/29/2025 the laboratory performed QC on lot number 0190 of DCA Vantage cartridges and receive an out of range QC result of 252.9 mg/L for albumin. 5. The DCA Vantage QC instrument tapes stated failed for the above out of range QC results. 6. TP #1 confirmed that the laboratory tested patients using the above lot numbers of DCA Vantage cartridges and did not take corrective action when QC fell outside of the established ranges.