

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 16D0383718	(X3) Date Survey Completed 02/06/2023
Name of Provider or Supplier Planned Parenthood Ames Health Center	Street Address, City, State 2530 Chamberlain St, Ames, IA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on review of proficiency records and reports, the laboratory fails to successfully participate in a proficiency testing program for the subspecialty, ABO group and D (Rho) typing, and analytes, ABO group and D(Rho) typing, for two consecutive proficiency testing events: 2022 testing events 2 and 3 (refer to D2162 and D2163).</p>
D2162	<p>ABO GROUP AND D(RHO) TYPING CFR(s): 493.859(f)</p>

Failure to achieve satisfactory performance for the same analyte in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:
Based on review of proficiency testing reports and records, the laboratory fails to achieve satisfactory performance for the analytes, ABO group and D(Rho) typing, for two consecutive testing events for unsuccessful participation. The laboratory received unsatisfactory performance scores of zero for 2022 testing event 2 and 80% for 2022 testing event 3.

D2163

ABO GROUP AND D(RHO) TYPING
CFR(s): 493.859(g)

Failure to achieve an overall testing event score of satisfactory for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:
Based on review of proficiency testing reports and records, the laboratory fails to achieve an overall testing event score of satisfactory performance for two consecutive testing events for the subspecialty, ABO group and D (Rho) typing. The laboratory received unsatisfactory performance scores of zero for 2022 testing event 2 and 80% for 2022 testing event 3 for the subspecialty, ABO group and D (Rho) typing.