

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 16D0385068	(X3) Date Survey Completed 07/02/2025
Name of Provider or Supplier Mercyone Kingsley Family Medicine Clinic	Street Address, City, State 111 Main Street, Kingsley, IA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A proficiency testing desk review survey was completed on July 2, 2025. The laboratory was found to be out of compliance with the following CONDITION LEVEL DEFICIENCIES: D2016 - 42 C.F.R 493.803 Condition: Successful Participation D6000 - 42 C.F.R 493.1403 Condition: Laboratory Director, moderate complexity
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on review of the Certification and Survey Provider Enhanced Reporting (CASPER) 155 report and graded results from the College of American Pathologists, the laboratory failed to successfully participate in two consecutive testing events for

	the analyte, hematocrit. The laboratory had unsatisfactory scores for 2025 event 1 and 2025 event 2. Refer to D2130.
D2130	HEMATOLOGY CFR(s): 493.851(f) (f) Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on review of the Certification and Survey Provider Enhanced Reporting (CASPER) 155 report and graded results from the College of American Pathologists (CAP), the laboratory failed to successfully participate in two consecutive testing events for the analyte, hematocrit. The findings include: 1. For 2025 event 1, the laboratory received an unsatisfactory performance score of 60% for the analyte, hematocrit. 2. For 2025 event 2, the laboratory receive an unsatisfactory performance score of 60% for the analyte, hematocrit. 3. The CASPER 155 report and graded results from CAP confirm the findings listed above.
D6000	MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403 The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart. This CONDITION is not met as evidenced by: Based on review of the Certification and Survey Provider Enhanced Reporting (CASPER) 155 report and graded results from the College of American Pathologists (2025 event 1 and 2025 event 2), the laboratory director failed to provide overall management and direction of the laboratory services. Refer to D6016.
D6016	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i) (e)(4)(i) The proficiency testing samples are tested as required under Subpart H of this part; This STANDARD is not met as evidenced by: Based on review of the Certification and Survey Provider Enhanced Reporting (CASPER) 155 report and graded results from the College of American Pathologists, the laboratory director failed to ensure successful participation in an HHS approved proficiency testing program. Refer to D2130.