

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  16D0385384	<b>(X3) Date Survey Completed</b>  02/07/2023
<b>Name of Provider or Supplier</b>  Sanford Sheldon Medical Center	<b>Street Address, City, State</b>  118 North Seventh Avenue, Sheldon, IA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5411</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(a)</p> <p>Test systems must be selected by the laboratory. The testing must be performed following the manufacturer's instructions and in a manner that provides test results within the laboratory's stated performance specifications for each test system as determined under 493.1253.</p> <p>This STANDARD is not met as evidenced by: Based on review of Stago coagulation test system records and confirmed by laboratory personnel identifiers #2 and #3 (refer to the Laboratory Personnel Report) at approximately 12:45 pm on 02/07/2023, the laboratory failed to verify the manual calculation of the international normalized ratio (INR) for one out of one lot number of prothrombin time reagent (259495, expiration 06/30/2023). The findings include: 1. At the time of the survey, the laboratory had in use prothrombin time reagent lot number 259495 (expiration 06/30/2023). 2. The coagulation reagent verification records for prothrombin time reagent lot number 259495 did not include verification of the accuracy of the INR calculation from the instrument. 3. Personnel identifiers #2 and #3 confirmed that the laboratory did not verify the accuracy of the INR calculation from the instrument.</p>