Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 16D0385556	**(X3) Date Survey Completed** 09/21/2022
Name of Provider or Supplier Lakes Regional Healthcare	**Street Address, City, State** 2301 Highway 71 South, Spirit Lake, IA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D6029	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(11) The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(11) Ensure that prior to testing patients' specimens, all personnel have the appropriate education and experience, receive the appropriate training for the type and complexity of the services offered, and have demonstrated that they can perform all testing operations reliably to provide and report accurate results. This STANDARD is not met as evidenced by: Based on review of personnel records, lack of training records and confirmed by laboratory personnel identifier #1 (refer to Laboratory Personnel Report) at approximately 10:00 am on 09/26/2022, the laboratory director failed to ensure that prior to testing patient specimens, testing personnel performing moderate complexity testing received the appropriate training for five out of five respiratory therapists (identifiers #2 - #6) The findings include: 1. In February 2022, the respiratory therapy staff, identifiers #2 - #6, started performing arterial blood gas testing on the i-STAT test system. 2. At the time of the survey, the laboratory did not have have training records available for respiratory therapy testing personnel identifiers #2 - #6.