

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 16D0386795	(X3) Date Survey Completed 06/27/2023
Name of Provider or Supplier Epiphany Dermatology Of Iowa, Plc	Street Address, City, State 5264 Council St Ne Ste 700, Cedar Rapids, IA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observations made during the survey, review of patient Mohs surgery logs and confirmed by laboratory personnel identifier #1 (refer to the Laboratory Personnel Report) at approximately 9:45 am on 6/27/23, the laboratory used expired tissue marking dye on 61 patients from 3/01/23 - 6/28/23. The findings include: 1. Laboratory personnel identifier #1 stated the laboratory used the red and blue tissue marking dyes on patient specimens when performing Mohs surgery. 2. The red tissue marking dye expired on 2/28/23 and the blue tissue marking dye expired on 3/31/23. 3. The laboratory performed Mohs surgery on 61 patients from 3/01/23 - 6/28/23.</p>