

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 16D0665914	(X3) Date Survey Completed 02/14/2022
Name of Provider or Supplier Eastern Iowa Health Center	Street Address, City, State 1201 3rd Avenue Se, Cedar Rapids, IA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5215	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2)</p> <p>The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results).</p> <p>This STANDARD is not met as evidenced by: Based on review of proficiency testing (PT) records and confirmed by laboratory personnel identifier #5 (refer to the Laboratory Personnel Report) at approximately 10:15 am on 02/14/2022, the laboratory failed to perform a self evaluation of ungraded PT scores for two out of six PT testing events from 01/01/2020- 12/31/2021 (2020 event 3 and 2021 event 3). The findings include: 1. For 2020 testing event 3, the laboratory received ungraded PT test scores for the following: *2020 Hematology /Coagulation 3rd event: urine sediment (specimen US-06) 2. For 2021 testing event 3, the laboratory received ungraded PT test scores for the following: *2021 Hematology /Coagulation 3rd event: vaginal wet prep (KOH) (Specimen VKP-03) 3. At the time of the survey, the laboratory did not have additional documentation or corrective action for the ungraded PT test scores.</p>