Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 16D0686166	**(X3) Date Survey Completed** 09/27/2024
Name of Provider or Supplier Clarinda Regional Health Center	**Street Address, City, State** 220 Essie Davison Drive, Clarinda, IA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D5215	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2) The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results). This STANDARD is not met as evidenced by: Based on review of proficiency testing (PT) records and confirmed by interview with General Supervisor (GS) identifier #2 at 9:41 am on 09/27/2024, the laboratory failed to perform a self evaluation when the laboratory received seven ungraded PT scores from three out of five testing events from 01/01/2023- 09/27/2024. The findings include: 1. For 2023 3rd event, the laboratory received ungraded PT test scores for the following: * Hematology/Coagulation: Polymorphonuclear leukocytes (PMN) (CSF /body fluid)- specimen BFL-04 * Hematology/Coagulation: Vaginal Wet Prep- specimen VA-03 2. For 2024 1st event, the laboratory received ungraded PT test scores for the following: * Hematology/Coagulation: Nucleated red blood cells (RBCs)- specimen COU-02 * Microbiology: Gram stain morphology- specimen GS-02 3. For 2024 2nd event, the laboratory received ungraded PT test scores for the following: * Hematology/Coagulation: Nucleated RBCs- specimen COU-06 * Hematology/Coagulation: Vaginal Wet Prep- specimen VA-02 * Microbiology: Gram stain morphology- specimen GS-08 4. At the time of the survey, GS #2 confirmed the laboratory did not perform a self evaluation for the ungraded PT test scores listed above.