

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 16D0875774	(X3) Date Survey Completed 02/11/2020
Name of Provider or Supplier Unitypoint Clinic Family Medicine - Johnston	Street Address, City, State 5900 Nw 86th Street, Suite 100, Johnston, IA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6029	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(11)</p> <p>The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(11) Ensure that prior to testing patients' specimens, all personnel have the appropriate education and experience, receive the appropriate training for the type and complexity of the services offered, and have demonstrated that they can perform all testing operations reliably to provide and report accurate results.</p> <p>This STANDARD is not met as evidenced by: Based on review of personnel records and confirmed by laboratory personnel identifier #1 (refer to the Laboratory Personnel Report) at approximately 3:15 pm on 02/11/2020, the laboratory director failed to ensure that prior to testing patient specimens, all testing personnel performing moderate complexity testing received the appropriate training for one out of one testing personnel (laboratory personnel identifier #4) who perform potassium hydroxide (KOH) preparations and wet mount examinations. At the time of the survey, the laboratory did not have training records available for personnel identifier #4.</p>