

<p>Statement of Deficiencies</p>	<p>(X1) Provider/Supplier/CLIA Identification Number</p> <p>16D0945914</p>	<p>(X3) Date Survey Completed</p> <p>03/07/2024</p>
<p>Name of Provider or Supplier</p> <p>James A Coggi, Md, Pc</p>	<p>Street Address, City, State</p> <p>1031 Office Park Road, Suite 10, West Des Moines, IA</p>	
<p>For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.</p>		

<p>(X4) ID Prefix Tag</p>	<p>Summary Statement of Deficiencies</p>
<p>D2128</p>	<p>HEMATOLOGY CFR(s): 493.851(e)</p> <p>(1) For any unsatisfactory analyte or test performance or testing event for reasons other than a failure to participate, the laboratory must undertake appropriate training and employ the technical assistance necessary to correct problems associated with a proficiency testing failure. (2) For any unacceptable analyte or testing event score, remedial action must be taken and documented, and the documentation must be maintained by the laboratory for two years from the date of participation in the proficiency testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of proficiency testing (PT) records and confirmed by laboratory personnel identifier #1 (refer to the Laboratory Personnel Report) at 9:45 am on 03/07/2024, the laboratory failed to take and document corrective action for seven unacceptable PT scores from three out of six PT testing events from 01/01/2022- 12/31/2023. The findings include: 1. For 2022 testing event 1, the laboratory received unacceptable PT test scores for the following: *2022 Hematology/Coagulation 1st event: Monocytes/Mids % (specimen HEM-04) 2. For 2022 testing event 2, the laboratory received unacceptable PT scores for the following: *2022 Hematology/Coagulation 2nd event: Lymphocytes % (specimen HEM-10) *2022 Hematology/Coagulation 2nd event: Monocytes/Mids % (specimen HEM-07 and HEM-10) 3. For 2023 testing event 3, the laboratory received unacceptable PT scores for the following: *2023 Hematology/Coagulation 3rd event: Lymphocytes % (specimen HEM-12) *2023 Hematology/Coagulation 3rd event: Monocytes/Mids % (specimen HEM-12 and HEM-13) 4. At the time of the survey, the laboratory did not have additional documentation or corrective action for the unacceptable PT test scores.</p>