

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 16D2056586	(X3) Date Survey Completed 08/13/2025
Name of Provider or Supplier Muscatine Fire Department	Street Address, City, State 312 East Fifth Street Ste 2, Muscatine, IA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6063	<p>LABORATORY TESTING PERSONNEL CFR(s): 493.1421</p> <p>The laboratory must have a sufficient number of individuals who meet the qualification requirements of 493.1423, to perform the functions specified in 493.1425 for the volume and complexity of tests performed.</p> <p>This CONDITION is not met as evidenced by: Based on review of personnel records and confirmed by interview with laboratory personnel identifier #1, the laboratory failed to ensure that individuals met the qualifications to perform moderate complexity testing as specified in 6065.</p>
D6065	<p>TESTING PERSONNEL QUALIFICATIONS CFR(s): 493.1423(b)(1)(2)(3)(4)(i)</p> <p>(b) Meet one of the following requirements: (b)(1) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; or (b)(2) Have earned a doctoral, master's, or bachelor's degree in a chemical, biological, clinical or medical laboratory science, or medical technology, or nursing from an accredited institution; or (b)(3) Meet the requirements in 493.1405(b)(3)(i)(B), (b)(4)(i)(B), (b)(4)(i)(C) or (b)(5)(i)(B); or (b)(4) Have earned an associate degree in a chemical, biological, clinical or medical laboratory science, or medical laboratory technology or nursing from an accredited institution; or (b)(5) Be a high school graduate or equivalent and have successfully completed an official military medical laboratory procedures course of at least a duration of 50 weeks and have held the military enlisted occupational specialty of Medical Laboratory Specialist (Laboratory Technician); or (b)(6)(i) Have earned a high school diploma or equivalent; and</p>

This STANDARD is not met as evidenced by:

Based on review of personnel records and confirmed by interview with personnel identifier #1 (refer to the Laboratory Personnel Report) at approximately 9:15 am on 8/13/2025, the laboratory failed to ensure two out of 23 testing personnel met the qualifications for moderate complexity testing personnel. The findings include: 1. Since the last survey performed on 4/6/2023, the laboratory trained testing personnel identifiers #2 and #3 to perform moderate complexity testing using the i-STAT test system. 2. Interview with personnel identifier #1 confirmed the laboratory did not have documents to qualify the above individuals to perform moderate complexity testing.