

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 16D2068702	(X3) Date Survey Completed 04/26/2023
Name of Provider or Supplier Advanced Dermatology Of The Midlands	Street Address, City, State 928 Valley View Drive, Council Bluffs, IA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observations made during the survey, review of patient test logs and confirmed by laboratory personnel identifier #1 (refer to the Laboratory Personnel Report) at approximately 1:30 pm on 4/26/2023, the laboratory failed to ensure that the potassium hydroxide (KOH) reagent did not exceed its' expiration date when used on four out of four patients from 2/1/2023 - 4/26/2023. The finding include: 1. At the time of the survey, the KOH reagent expired on 1/31/2023. 2. The laboratory had performed KOH preparations on four patients from 2/1/2023 - 4/26/2023.</p>
D5805	<p>TEST REPORT CFR(s): 493.1291(c)</p> <p>The test report must indicate the following: (c)(1) For positive patient identification, either the patient's name and identification number, or a unique patient identifier and identification number. (c)(2) The name and address of the laboratory location where the test was performed. (c)(3) The test report date. (c)(4) The test performed. (c)(5) Specimen source, when appropriate. (c)(6) The test result and, if applicable, the units of measurement or interpretation, or both. (c)(7) Any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.</p> <p>This STANDARD is not met as evidenced by:</p>

Based on review of patient test reports and Mohs maps and confirmed by laboratory personnel #1 (refer to the Laboratory Personnel Report) at approximately 1:30 pm on 4/26/2023, the laboratory failed to include the address of the testing facility on three out of five patient test reports reviewed from 1/1/2023 - 4/26/2023. The findings include: 1. The laboratory performed Mohs surgery, including the reading of Mohs slides, on 1/18/2023, 1/23/2023, and 1/26/2023. 2. The Mohs maps from each of the dates confirmed that the provider read the Mohs slides at Advanced Dermatology of the Midlands - Council Bluffs site. 3. At the time of the survey, the patient test reports did not have the correct address of the testing facility.