

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 16D2131355	(X3) Date Survey Completed 10/14/2022
Name of Provider or Supplier Vascular Institute Of The Midwest	Street Address, City, State 3385 Dexter Court Pavillion 3, Suite 100, Davenport, IA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on review of proficiency testing records and reports, the laboratory failed to successfully participate in a proficiency testing program for the specialty, hematology, for two consecutive proficiency testing events: 2022 events 1 and 2 (refer to D2131).
D2131	HEMATOLOGY CFR(s): 493.851(g) Failure to achieve an overall testing event score of satisfactory performance for two

consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on review of proficiency testing reports and records, the laboratory failed to achieve an overall testing event score of satisfactory performance for two consecutive testing events for the specialty, hematology; the laboratory received unsatisfactory performance scores of 60% for 2022 testing event 1 and 33% for 2022 testing event 2. The findings include: 1. The laboratory received an unsatisfactory performance score of zero for 2022 testing event 1 for the analyte, prothrombin time. 2. The laboratory received unsatisfactory performance scores of zero for 2022 testing event 2 for the analytes, hematocrit and hemoglobin.