Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	16D2131355	11/20/2025
Name of Provider or Supplier	**Street Address, City, State**	
Vascular Institute Of The Midwest	3385 Dexter Court Pavillion 3, Suite 100, Davenport, IA	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5431	MAINTENANCE AND FUNCTION CHECKS CFR(s): 493.1254(a)(2) (a)(2) Function checks as defined by the manufacturer and with at least the frequency specified by the manufacturer. Function checks must be within the manufacturers established limits before patient testing is conducted. (b) Equipment, instruments, or test systems developed in-house, commercially available and modified by the laboratory, or maintenance and function check protocols are not provided by the manufacturer. The laboratory must do the following: This STANDARD is not met as evidenced by: Based on review of i-Stat thermal probe records, review of the i-Stat operator's guide, and confirmed by interview with Testing Personnel #1 (TP #1) at 9:45 am on 11/20 /2025, the laboratory failed to perform and document thermal probe function checks on the i-Stat analyzer every six months for one out of four time periods from 01/01 /2024- 11/20/2025. The findings include: 1. The laboratory used an i-Stat test system to perform the following tests: protime, activated clotting time (ACT), pH, partial pressure of carbon dioxide (PCO2), partial pressure of oxygen (PO2), lactate, sodium, potassium, chloride, carbon dioxide, glucose, ionized calcium, blood urea nitrogen, creatinine, and hematocrit. 2. The i-Stat operator's guide stated that a thermal probe check must be performed every six months. 3. Review of the laboratory's i-Stat records indicated that the laboratory last performed a thermal probe check on 06/18 /2024. 4. At the time of the survey, TP #1 confirmed the laboratory did not perform thermal probe checks at any other time from 01/01/2024- 11/20/2025.