

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 16D2171667	(X3) Date Survey Completed 04/28/2022
Name of Provider or Supplier Cornerstone Pediatrics & Family Allergy	Street Address, City, State 6800 Lake Drive Suite 260, West Des Moines, IA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2128	<p>HEMATOLOGY CFR(s): 493.851(e)</p> <p>(1) For any unsatisfactory analyte or test performance or testing event for reasons other than a failure to participate, the laboratory must undertake appropriate training and employ the technical assistance necessary to correct problems associated with a proficiency testing failure. (2) For any unacceptable analyte or testing event score, remedial action must be taken and documented, and the documentation must be maintained by the laboratory for two years from the date of participation in the proficiency testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of proficiency testing (PT) records and confirmed by laboratory personnel identifier #2 (refer to the Laboratory Personnel Report) at approximately 1: 45 pm on 04/28/2022, the laboratory failed to take and document corrective action for eight unacceptable PT scores from two out of seven PT testing events (2021 events 1 and 3) from 01/01/2020- 04/28/2022. The findings include: 1. For 2021 testing event 1, the laboratory received unacceptable PT test scores for the following: *2021 Hematology/Coagulation 1st event: Mean Corpuscular Hemoglobin (MCH) (specimen HSY-01) *2021 Hematology/Coagulation 1st event: Platelet Count (specimens HSY-01 and HSY-02) *2021 Hematology/Coagulation 1st event: Red Cell Count (specimen HSY-01) 2. For 2021 testing event 3, the laboratory received unacceptable PT scores for the following: *2021 Hematology/Coagulation 3rd event: MCH (specimen HSY-11) *2021 Hematology/Coagulation 3rd event: Mean Platelet Volume (MPV) (specimen HSY-11) *2021 Hematology/Coagulation 3rd event: Platelet Count (specimen HSY-11) *2021 Hematology/Coagulation 3rd event: Red Cell Count (specimen HSY-11) 3. At the time of the survey, the laboratory did not have additional documentation or corrective action for the unacceptable PT test scores.</p>