

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 16D2307836	(X3) Date Survey Completed 08/19/2026
Name of Provider or Supplier Uph Emergency Department-Marion	Street Address, City, State 3301 Armar Dr, Marion, IA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5215	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2) The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results). This STANDARD is not met as evidenced by: Based on review of proficiency testing (PT) records and confirmed by interview with Technical Supervisor #1 (TC #1) at 9:25 am on 8/19/2026, the laboratory failed to perform a self evaluation when the laboratory received ungraded PT scores for four out of five testing events from 01/01/2025- 8/19/2026. The findings include: 1. For 2025 event 1, the laboratory received an ungraded PT score for albumin, specimen CH-04. 2. For 2025 event 2, the laboratory received ungraded PT scores for total bilirubin: specimens CH-07, CH-09 and CH-10. 3. For 2025 event 3, the laboratory received ungraded PT scores for total bilirubin: specimens CH-12 and CH-15. 4. For 2026 event 1, the laboratory received ungraded PT scores for albumin, specimen CH-05 and total bilirubin, specimen CH-05. 2. At the time of the survey, TC #1 confirmed the laboratory did not perform a self evaluation for the ungraded PT test scores listed above.