

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 17D0046900	(X3) Date Survey Completed 07/05/2023
Name of Provider or Supplier Wilson Medical Center	Street Address, City, State 2600 Ottawa Road, Neodesha, KS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on the review of proficiency testing (PT) records from American Proficiency Institute (API), the laboratory failed to successfully participate in PT under the specialty Endocrinology for the regulated analyte: 0585 Thyroid-Stimulating Hormone (TSH) for two out of three consecutive testing events. (Refer to D2107)
D2107	ENDOCRINOLOGY CFR(s): 493.843(f) Failure to achieve satisfactory performance for the same analyte or test in two

consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on a desk review of proficiency testing (PT) records from the CMS report 0155D and phone interview, the laboratory failed to achieve satisfactory performance in PT for the regulated analyte 0585 Thyroid-Stimulating Hormone (TSH) for two out of three consecutive testing events. Findings: 1. A review of proficiency testing records from the PT provider American Proficiency Institute (API) Proficiency Testing for the regulated analyte 0585 TSH revealed the following performance scores for the periods indicated: 2. Third testing event, 2022 revealed a score 0% 3. Second testing event, 2023 revealed a score of 40% 4. Phone interview on July 5 2023 at 3:00 p.m. confirmed, the laboratory failed to achieve satisfactory performance in PT for the regulated analyte 0585 TSH for two out of three consecutive testing events.