

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 17D0449875	(X3) Date Survey Completed 01/19/2022
Name of Provider or Supplier Associated Urologists, Pa	Street Address, City, State 1133 College Ave, G100, Manhattan, KS	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6046	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(8)</p> <p>(b) The technical consultant is responsible for-- (b)(8) Evaluating the competency of all testing personnel and assuring that the staff maintain their competency to perform test procedures and report test results promptly, accurately and proficiently.</p> <p>This STANDARD is not met as evidenced by: Based on the lack of documentation and interview with the laboratory director (LD), the technical consultant failed to evaluate and document competency on five of five testing personnel for all of 2020, 2021 and to date 1/19/22. Findings: 1. Request was made for competency documentation for all testing personnel. No documentation of competency assessment for five of five testing personnel was made available at the time of survey. 2. Interview with the LD on 1/19/22 at 1:40 p.m. confirmed, the technical consultant failed to evaluate and document competency on five of five testing personnel for all of 2020, 2021 and to date 1/19/2022.</p>